# NSN 4920-00-944-8357

Gyro Test Station - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/4920-00-944-8357

### **Functional Classification:**

Aa-9.6

### **Special Features:**

C/o signal control panel; dc null detector and vtvm; servo amplifier; torquer bias panel; oscilloscope; six-channel recorder; servo amplifier power supply; meter panel; temp recorder; temp control panel; pulsed 30 wheel supply; power panel; oscillator; counter; counter signal panel; power control panel; wattmeter; filament supply; 250.0 vdc power supply; 300.0 vdc power supply; power junction box; pedestal; blower; heater; gimbal assembly; optical block; mirror; cable assemblies; 115.0 vac, 60.0-400.0 hz, single/three phase

## **Functional Description:**

Used to check axis alignment, measure/check scale factor and torquer resistance of the gyro under test

### **Reference Data And Literature:**

T.O. 33d7-63-5-1

Shelf Life:

N/a

**Unit Of Measure:** 

--

**Demilitarization:** 

No

Fiig:

A23800