

View Online at <https://aerobasegroup.com/nsn/4940-01-473-2540>

**End Application:**

F-16a/b/c/d

**Usage Design:**

Bused to verify the accuracy of the acceptance test stand, regulator tester (p/n 3301332-6101) and the pump test stand p/n 3301331-6101 as developed by litton instrument and life support division

**Product Name:**

Mass flow calibration system

**Special Features:**

C/o molblox model 1 part number fam0004, dual-mid/high flow molstic model 401-317, 0-10 slpm molblox model 401-080 0-40slpm molblox 401-042 and outlet-tee model 122581 Or equivalent; the range of the mass flow calibration system is 0.0 to 40 slpm, with an accuracy of the greater of, +/- 0.2% of thereading, or 0.02 of fs

**Reference Data And Literature:**

F42620-97-d-0010, serd 47048

**Supplementary Features:**

Document number 16pr011

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

Yes - demil/mli

**Fiig:**

A23800