

View Online at <https://aerobasegroup.com/nsn/5945-01-224-0109>

**Overall Length:**

2.500 inches

**Center To Center Distance Between Mounting Facilities Parallel To Length:**

0.200 inches single mounting facility

**Center To Center Distance Between Mounting Facilities Parallel To Width:**

0.700 inches single mounting facility

**Overall Height:**

1.400 inches

**Overall Width:**

1.400 inches

**Case Material:**

Plastic

**Unthreaded Mounting Hole Diameter:**

0.168 inches single mounting facility and 0.174 inches single mounting facility

**Contact Material:**

Copper

**End Application:**

6525-01-159-3009; end item name: radiographic/fluoroscopic x-ray system model dxd 350 ii

**Frequency In Hertz:**

Between 50.0 and 60.0

**Purpose For Which Designed:**

Specific

**Coil Quantity:**

1

**Duty Type:**

Continuous

**Mounting Facility Type And Quantity:**

1 unthreaded hole single mounting facility

**Pile-up Main Contact Form Arrangement:**

2c

**Cubic Measure:**

4.900 cubic inches

**Operating Current Rating And Type:**

5.0 amperes ac all windings and 3.0 amperes ac all windings

**Main Contact Load Current Rating At Maximum Rated Voltage:**

3.0 amperes ac resistive load

**Main Contact Maximum Voltage Rating In Volts:**

250.0 ac

**Precious Material And Location:**

Cross-bar gold

**Precious Material And Weight:**

0.013 gold ounces, troy

**Precious Material:**

Gold

**Terminal Type And Quantity:**

11 tab, solder lug all functions

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A03300