

View Online at <https://aerobasegroup.com/nsn/6515-01-277-6380>

End Application:

6515-01-185-8446

Product Name:

Vaporizer anesthetic u/w anes appar-oxygen monitor mdl 5120

General Description:

Ivaporizer thermometer p/n 205-8811-300, (located at bottom side of vaporizer) is provided to show the liquid anesthetic agent temperature within the vaporizing chamber, dial indicator 0-50 deg. C; the anesthetic agent saturation ratio is the amount of vehicle oxygen; divided by the amount of agent vapor in the vehicle oxygen; each agents saturation ratio is determined by temperature and pressure; for use with anesthesia apparatus-oxygen monitor, mdl. 5120

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No