

View Online at <https://aerobasegroup.com/nsn/6515-01-501-2309>

**End Application:**

Electrosurgery unit

**Product Name:**

Electrode indifferent electromedical variable tip loop 2s

**Special Features:**

Electrode loop

**Supplementary Features:**

All macan accesories may be sterilized by standard methods, all should be clean of blood and tissue debris before sterilization. Standard methods of sterilization include: autoclave: 273 f at 30 lbs pressure. Dry heat: 320 f, 1 hour. Chemical: any bactericide.

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A23900