

View Online at <https://aerobasegroup.com/nsn/6515-01-501-2319>

End Application:

Electrosurgery unit

Product Name:

Electrode indifferent electromedical var tip loop 2s

Special Features:

Electrode loop

Supplementary Features:

All macan accesories may be sterilized by standard methods, all should be clean of blood and tissue debris before sterilization. Standard methods of sterilization include: autoclave: 273 f at 30 lbs pressure. Dry heat: 320 f, 1 hour. Chemical: any bactericide.

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23900