# NSN 6515-01-506-2383

X-ray Generator - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6515-01-506-2383

#### Wattage Rating:

50.000 kilowatts

### Voltage In Volts And Current Type:

240.0 ac

#### **Special Features:**

Microprocessor control; apr operating mode with 100 examination distributed across 10 anatomical regions; manual 3-factor selection to modify the recommended kv, ma and time; auto-calibration; automatic tube protection; 30-80 kw; 40-150 kvp kv range; 20-1000 ma range; 0.4-600mas range; 3 ms-1000 sec exposure time; 40-120 fluoroscopic kv; 0.5-10 fluoroscopic ma **Shelf Life:** 

-----

N/a

## Unit Of Measure:

---

#### Demilitarization:

No