## NSN 6515-01-573-5373

Anesthesia Equipment Heat And Moisture Exchanger Filter - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6515-01-573-5373

## **End Application:**

Used on edith flex heat and moisture exchanger for intubation anesthesia tracheotomy patients

## **Special Features:**

Y filter/s plus edith flex; 30 mg oxygen moisture; removes 99.9 pct of bacteria/virus; 15f and 2 x 22m connectors; pkg of 20; in operating room, change after ea patient; in icu, change after 24 hrs

Shelf Life:

N/a

**Unit Of Measure:** 

Demilitarization:

No

Fiig:

T117-a