## NSN 6520-01-532-8136

Fiig: T243-n

Dental Laboratory Grinding And Polishing Machine - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6520-01-532-8136 **Current Type:** Ac Frequency Rating: 60.0 hertz Voltage Rating: 115.0 volts **Power Source:** Electric motor **Mounting Type For Which Designed:** Bench **Spindle End Type:** Double Spindle Speed In Rpm: 1725.0 and 3450.0 **Product Name:** Chuk/changer with lathe Phase: Single **Special Features:** Factory installed 16 chuk/changer to the right side of 26a low base lathe provides chuck changing efficiency at both low and high lathe speeds; movement of the activating handle allows technician to change from 1/4" to 3/32" burs, mandrels and chuck accessories without stopping the lathe Shelf Life: N/a **Unit Of Measure: Demilitarization:** No