

View Online at <https://aerobasegroup.com/nsn/6525-00-027-6593>

Peak Current Rating:

5.0 microamperes

Tube Insert Anode Type:

Stationary

Tube Insert Fluoroscopic Peak Voltage Rating:

85.0 kilovolts

Tube Insert Radiographic Peak Voltage Rating:

90.0 kilovolts

Radiographic Peak Current Rating:

100.0 microamperes

Radiographic Exposure Time Rating:

2.0 seconds

Focal Spot Quantity:

1 and 1

Focal Spot Size:

1.5 mm sq; 2.3 mm sq

Product Name:

Tube insert medical x-ray app radio& fluoro stat anode100ma90pkv

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T140-c