NSN 6525-00-222-1340

Dental X-ray Apparatus - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6525-00-222-1340

| Mounting Type: |
|---|
| Wall |
| |
| Accessory Components And Quantity: |
| Cassette 1; film processing hangars 3; head measuring calipers 1; intra-oral midline locators 2 |
| Current Rating In Amps: |
| 10.0 |
| Design Type: |
| Stationary |
| Operation Method: |
| Automatic |
| Control Type: |
| Cabinet |
| Tube Quantity: |
| 1 |
| Electric Power Rating: |
| 2400.0 watts |
| Tube Type: |
| Line focus |
| Rectification Type: |
| Half-wave |
| Tube Cooling Method: |
| Oil |
| Tube Mounting Type: |
| Tube stand |
| Tube Stand Mounting Type: |
| Base |
| Radiographic Timer Type: |
| Exposure |
| Transformer Rectification Method: |
| Self-rectifying |
| Transformer Output Capacity: |
| 10.0 output ma at 90 kv peak |
| Voltage In Volts: |
| 110.0 and 125.0 |
| Phase: |
| Single |
| Special Features: |
| FORR-EP-LAC-EMEN |
| Frequency In Hertz: |
| 50.0 or 60.0 |

Supplementary Features:

Patients head concentrentic positioner; panographic; motor driven chair

NSN 6525-00-222-1340

Dental X-ray Apparatus - Page 2 of 2



| _ | | | | - | | |
|---|---|---|-----|---|------|--|
| C | h | 0 | Iŧ. | | ife: | |
| | | | | | | |

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T140-c