## NSN 6525-00-222-1340

Dental X-ray Apparatus - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6525-00-222-1340

Mounting Type:
Wall
Accessory Components And Quantity:
Cassette 1; film processing hangars 3; head measuring calipers 1; intra-oral midline locators 2
Current Rating In Amps:
10.0
Design Type:
Stationary
Operation Method:
Automatic
Control Type:
Cabinet
Tube Quantity:
1
Electric Power Rating:
2400.0 watts
Tube Type:
Line focus
Rectification Type:
Half-wave
Tube Cooling Method:
Oil
Tube Mounting Type:
Tube stand
Tube Stand Mounting Type:
Base
Radiographic Timer Type:
Exposure
Transformer Rectification Method:
Self-rectifying
Transformer Output Capacity:
10.0 output ma at 90 kv peak
Voltage In Volts:
110.0 and 125.0
Phase:
Single
Special Features:
FORR-EP-LAC-EMEN
Frequency In Hertz:
50.0 or 60.0

Supplementary Features:

Patients head concentrentic positioner; panographic; motor driven chair

## NSN 6525-00-222-1340

Dental X-ray Apparatus - Page 2 of 2



_				-		
C	h	0	Iŧ.		ife:	

N/a

**Unit Of Measure:** 

--

Demilitarization:

No

Fiig:

T140-c