NSN 6525-01-013-3613

No Fiig: T140-c

Medical X-ray Apparatus Tube Unit - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6525-01-013-3613 **Container Type:** Any acceptable **Features Provided:** Grid control **Tube Insert Anode Type:** High speed rotating **Tube Insert Radiographic Peak Voltage Rating:** 150.0 kilovolts **Focal Spot Quantity: Focal Spot Size:** 0.3 mm sq, 1.0 mm sq **Housing X-ray Protective Value:** X-ray protective value of housing shall be in accordance with dhew publication no. Fda 75-8003 jul 1974 **Target Base Material:** Molybdenum **Target Angle In Deg:** 10.0 **Target Face Material:** Tungsten-rhenium alloy Port To Cable Receptacle Angle In Deg: 270.0 **Anode Heat Capacity Rating In Heat Units:** 300000 **Product Name:** Tube unit medical x-ray apparatus focal spot size 0.3; 1.0mm sq Shelf Life: N/a **Unit Of Measure: Demilitarization:**