

View Online at <https://aerobasegroup.com/nsn/6525-01-020-8353>

**Container Type:**

Any acceptable

**Tube Insert Anode Type:**

High speed rotating

**Tube Insert Radiographic Peak Voltage Rating:**

150.0 kilovolts

**Focal Spot Quantity:**

2

**Focal Spot Size:**

0.6 mm sq; 1.5 mm sq

**Housing X-ray Protective Value:**

X-ray protective value of housing shall be in accordance with dhew publication no. Fda 75-8003 jul 1974

**Target Base Material:**

Molybdenum

**Target Angle In Deg:**

11.0

**Target Face Material:**

Tungsten-rhenium alloy

**Port To Cable Receptacle Angle In Deg:**

180.0

**Anode Heat Capacity Rating In Heat Units:**

350000

**Product Name:**

Tube unit medical x-ray apparatus hi-speed rotating anode 150pkv

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

T140-c