

View Online at <https://aerobasegroup.com/nsn/6525-01-130-1389>

Cooling Method:

Oil

Rectification Type:

Half-wave

Tube Insert Anode Type:

Stationary

Tube Insert Radiographic Peak Voltage Rating:

70.0 kilovolts

Err Default:

Err default

Err Default:

Err default or err default

Focal Spot Quantity:

1

Transformer Rectification Method:

Electron tube

Err Default:

Err default

Product Name:

Tube-transformer head dental x-ray apparatus 70 kilovolts

Special Features:

Includes long cone and suspension arm

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fig:

T140-c