

View Online at <https://aerobasegroup.com/nsn/6525-01-207-0824>

Mounting Type:

Portable

Current Rating In Amps:

10.0

Design Type:

Portable

Operation Method:

Manual

Control Type:

Console

Tube Quantity:

1

Tube Type:

Single focus

Rectification Type:

Half-wave

Tube Cooling Method:

Oil

Tube Mounting Type:

Extension bracket

Tube Stand Mounting Type:

Tripod in case

Radiographic Timer Type:

Hand

Transformer Rectification Method:

Self-rectifying

Transformer Output Capacity:

7 ma at 70 kv peak

Product Name:

X-ray apparatus field dental:

Voltage In Volts:

110.0 and 220.0

Phase:

Single

Special Features:

The x-ray unit and collapsible patient seat shall be in rugged metal container, made of 6 handles. The handles shall be 2 per side, on opposite sides, and one per end panel, on opposite ends. Container will have permanently attached gasket seal (butyl rubber) resist entry from rain and dust. Case will be cushioned by high-density unicellular polyethylene foam. Case will contain a pressure relief valve. Unit mounted on tripod in the carrying case. Leveling method shall be employed. The case leveling bracket receptacles will be reinforced to protect against transport damages.

Frequency In Hertz:

50.0 or 60.0

Shelf Life:

N/a

Unit Of Measure:

Demilitarization:

No

Fig:

T140-c