

View Online at <https://aerobasegroup.com/nsn/6525-01-216-3378>

**Accessory Components And Quantity:**

Skin spacer, 1

**Design Type:**

Mobile

**Tube Quantity:**

1

**Camera Type:**

High resolution, ultra-low light level, tv

**Tube Type:**

Double focus, rotating anode radiographic and fluroscopic

**Rectification Type:**

Full-wave

**Tube Mounting Type:**

Movable c arm and movable l arm

**Tube Stand Mounting Type:**

Base

**Transformer Rectification Method:**

Solid state

**Transformer Output Capacity:**

30 ma at 100 kvp

**Err Default:**

Err default

**Product Name:**

X-ray apparatus radiographic and fluoroscopic: c-arm

**Voltage In Volts:**

115.0 and 230.0

**Phase:**

Single

**Special Features:**

Manual reversal up to 180 degrees w/o turning l assembly upside down, manual orbital movement of arm arc 90 degrees min, arm assembly motorized rotation 360 deg, horizontal position adjustable 7 in. Max, vertical position adj 17 in. Max, wig-wag movement plus or minus 12.5 deg; video disk recorder and two 14 in. Tv monitors on wheeled mounted base

**Frequency In Hertz:**

50.0 or 60.0

**Supplementary Features:**

Solid state; tube mounting c and l arm; image intensification system; beam limiting device; skin spacer; 500 watts standby, 1200 watts max

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

T140-c