

View Online at <https://aerobasegroup.com/nsn/6525-01-290-4687>

**Drive Type:**

Motor

**Accessory Components And Quantity:**

Adjustable non-skid footrest 1; compression-immobilization device 1; head clamp 1; patient grasp handles 1 pr; skidproof patient step 1; shoulder rest 1; urological knee crutches 1 pr

**Design Type:**

Stationary

**Tube Quantity:**

2

**Tube Type:**

Double focus, rotating anode fluoroscopic and double focus, rotating anode radiographic

**Table Type:**

Tilting

**Rectification Type:**

Full-wave

**Tube Cooling Method:**

Air

**Tube Mounting Type:**

Table and tube stand

**Tube Stand Mounting Type:**

Floor

**Err Default:**

Err default

**Err Default:**

Err default

**Transformer Rectification Method:**

Self-rectifying

**Transformer Output Capacity:**

To 300 ma at 150 kv peak

**Product Name:**

Xray apparatus radiographic double focus 2 tube 115/120v50/60hz

**Voltage In Volts:**

115.0 and 120.0

**Phase:**

Single

**Frequency In Hertz:**

50.0 or 60.0

**Departure From Cited Document:**

Peacetime demand obtain

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

T140-c