NSN 6525-01-509-0515

Demilitarization:

No Fiig: T140-c

Dental X-ray Apparatus - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6525-01-509-0515 **Mounting Type:** Stationary **Current Rating In Amps:** 13.0 **Design Type:** Stationary **Operation Method:** Automatic **Control Type:** Vertical panel Film Width: Between 6.000 inches and 12.000 inches **Err Default:** Err default **Err Default:** Err default **Transformer Output Capacity:** 60.0 to 80 kv, in2 kv steps Err Default: Err default **Product Name:** Hight tech, software controlled digital compatable wheelchair accessibile paniceph Voltage In Volts: 120.0 and 230.0 **Special Features:** Trolled, motion path, three pointpositioning lights, multiple imaging programs Frequency In Hertz: 50.0 or 60.0 Shelf Life: N/a **Unit Of Measure:**