NSN 6530-00-710-5512

Diathermy Apparatus - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6530-00-710-5512

Accessory Components And Quantity:

Applicator, air spaced, 1; applicator, cable, 1; applicator, condenser pad, 4 by 18 in., approx, 2; applicator, condenser pad, 7 by 9 in., le,

approx, 2; applicator, drum, hinged type, 1; arm, supporting, adjustable, 2; bandage, 2; cable, line supply, 1; cord, connecting, orificial
electrode, 45 in. Lg, approx, 1; electrode, ball, 3/16 in., 1; electrode, curved point, coagulating, 1; electrode, needle, 7/8 in. Shank, 1;
electrode, prostatic, orificial, 1; electrode, round loop, 1/4 in., 7/8 in. Shank, 1; electrode, round loop, 1/4 in., 2-3/4 in. Shank, 1; electrode,
round loop, 5/16 in., 2-3/4 in. Shank, 1; electrode, straight point, coagulating, 1; electrode, vaginal, orificial, 1; fuse, spare, 2; guard, cable,
1; handle, chuck, w/make and break switch, 1; needle, coarse, 2-3/4 in. Shank, 1; separator, clip, for calbr applicator, 6; set instructions,
operating, incl the wiring diagram, 1; spacer, felt, 4 by 18 in., approx, 6; spacer, felt, 7 by 9 by 3/8 in., approx, 6
Installation Design:
Mobile
Input Current Rating:
15.0 amperes
Output Wattage Rating:
200.0 watts
Frequency Rating In Megahertz:
27120.0
Control Name And Quantity:
Compensator, line voltage, 1; control, output, 1; indicator, filament voltage, 1; indicator, power output, 1; light, pilot or siutable indicator, 1;
switch, on and off, 1
Voltage In Volts:
105.0 or 125.0
Frequency In Hertz:
60.0
Specification Or Standard:
C class
Departure From Cited Designator:
As modified by control name and quantity and accessory components and quantity
Shelf Life:
N/a
Unit Of Measure:
Demilitarization:
No
Fiig:
T107-c