

View Online at <https://aerobasegroup.com/nsn/6530-01-510-4685>

**Product Name:**

Attest rapid readout biological indicator

**General Description:**

7reads fluorescene produced by positive 3m attest rapid readout biological indicators 1291 and 1292 to indicate a sterilization process; audible alarm and red light indicates sterilization failure; green light indicates sterilization conditions met

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No