

View Online at <https://aerobasegroup.com/nsn/6530-01-529-4093>

**End Application:**

Attest incubator models 126, 130, and 116

**Physical Form:**

Vial

**Product Name:**

3m attest biological indicators

**Special Features:**

Used with 270.0 degree f. - 132.0 degree c. Gravity steam flash sterilizers; includes bacillus stearothermophilus spore strip; sealed glass ampule with growth medium and a bromocresol purple ph indicator dye system; blue color-coded cap; hydrophobic filter; chemical indicator on label changes from rose color to brown color when processed; 24 hour incubation period; yellow color indicates positive result

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

T396-a