# NSN 6550-01-500-2390

Gram Positive Pa Culture Control - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6550-01-500-2390

## **End Application:**

For use in performance testing of culture media, stains, biochemical id kits, maintenance of stock cultures and in the evaluation of bacteriological procedures

### **Product Name:**

Cutlure control gram positive panel microscan culti loop set 5s

### **Special Features:**

Microscan culti-loop set; to grow gram positive culture control organisms for quality control of microscan instruments; contains stabilized, preserved viable microorganisms; each loop may be dissolved in re-hydration media or plated/streaked directly onto appropriate media; includes: (5) enterococcus faecalis atcc 29212, (5) micrococcus luteus atcc 49732, (5) staphylococcus aureus atcc 29213, and (5) streptococcus bovis atcc 49147; should be stored upon arrival at 2 to 8 degrees c (or frozen for campylobacter sp.) each loop is individually packaged in a foil pouch with a peel-off label

Ch.	_lf	Life:	
OII	eII	LIIE.	

N/a

**Unit Of Measure:** 

--

#### **Demilitarization:**

No