# NSN 6550-01-519-3894

Activated Partial T Test Cuvette - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6550-01-519-3894

## **Product Name:**

Aptt testing cuvettes for use with nsn 6630-01-519-3589 coagulation analyzer box 45s

## **Special Features:**

Activated partial thromboplastin time test cuvette for use with fresh whole blood; use with nsn 6630-01-519-3589 coagulation analyzer; box

of 45 cuvettes. (do not freeze), refrigerat between 2-8 degrees celsius or 35-46 degree fahrenheit.

## Supplementary Features:

For use with hemochron signature jr

## Shelf Life:

N/a

Unit Of Measure:

---

## Demilitarization:

No