

View Online at <https://aerobasegroup.com/nsn/6550-01-519-5571>

**Primary Container Type:**

Vial

**Product Name:**

Abnormal level plasma control for citrated aptt use w/coagulation analyzer nsn 6630-01-519-3589 hemochron jr. Citrate aptt cuvettes only box includes 15 abnormal vials and 4 protective sleeves

**Special Features:**

Microcoagulation quality control direct check; abnormal level plasma control for citrated aptt; use w/coagulation analyzer, nsn 6630-01-519-3589; each box includes 15 normal level vials and 4 protective sleeves; use as an abnormal level control solution for citrated samples for aptt testing only; for citrated samples only; abnormal level qc for use with citrate aptt cuvettes only

**Supplementary Features:**

For use with hemochron jr. Citrate aptt cuvettes only

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A10000