

View Online at <https://aerobasegroup.com/nsn/6550-01-527-7062>

Special Features:

Calibration stability is 56 days; serum and plasma (heparin)/20ul; 74 min. Time to first result; 2.2 iu/ml analytical sensitivity; 0, 50, 250, 500, 1000 and 2500 iu/ml calibrator levels; 56 days open pack stability; consists of tri (hydroxymethyl) aminomethane and tris (hydroxymethyl) aminomethane hydrochloride

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No