NSN 6625-00-079-3676

Oscilloscope - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6625-00-079-3676

| view Offilite at https://aerobasegroup.com/hish/oo25-oo-o/5-50/0 |
|---|
| Current Type: |
| Ac |
| Frequency Rating: |
| Between 50.0 hertz and 400.0 hertz |
| Accessory Components And Quantity: |
| Adapter 2, cable assembly 1, cord, power 1, filter, light 1, lead, ground 1, lead, test 1, probe, test 1, tip, test 1 |
| Voltage Rating: |
| Between 105.0 volts and 125.0 volts |
| Beam Quantity: |
| 1 |
| Accuracy Range In Percent: |
| +0.0 to 3.0 |
| Screen Size: |
| 5 in. |
| Expanded Sweep: |
| Provided |
| Amplifier Deflection Sensitivity Provision: |
| Not provided |
| Input Intensity Modulation Feature: |
| Not provided |
| Plate Direct Connection Deflection Sensitivity Provision: |
| Not provided |
| Frequency Response Rating: |
| Provided |
| X-axis Frequency Response Rating: |
| Between 0.0 megahertz and 15.0 megahertz |
| Amplifier Input Impedance Rating Provision: |
| Provided |
| X-axis Input Impedance Rating: |
| 1.0 megohms |
| Y-axis Input Impedance Rating: |
| 1.0 megohms |
| Amplitude Calibrating Voltage Waveform: |
| Square wave |
| Calibrating Peak To Peak Amplitude: |
| 1.0 volts |
| Measurement Range Quantity: |
| 3 |
| Measurement Range Steps: |
| 0.05 to 20 v per cm |
| Calibrated Screen Scale: |

Not included

NSN 6625-00-079-3676

Oscilloscope - Page 2 of 2



| Trace Storage Tube: |
|---|
| Not included |
| Camera Recording Device Accommodation: |
| Not provided |
| Phase: |
| Single |
| Fsc Application Data: |
| 1. Oscilloscope, except, specially designed |
| Shelf Life: |
| N/a |
| Unit Of Measure: |
| |
| Demilitarization: |
| No |
| Fiig: |

T320-g