

View Online at <https://aerobasegroup.com/nsn/6625-00-257-1582>

**Circuit Current For Which Designed:**

Ac

**Reliability Indicator:**

Not established

**Overall Length:**

2.016 inches

**Center To Center Distance Between Mounting Facilities Parallel To Length:**

4.000 inches

**Center To Center Distance Between Mounting Facilities Parallel To Width:**

3.562 inches

**Overall Height:**

4.260 inches

**Overall Width:**

5.620 inches

**Mounting Stud Length:**

0.380 inches

**Furnished Items And Quantity:**

Thermocouple w/4 feet leads 1

**Installation Design:**

Panel

**Mounting Arrangement Style:**

Four hole/stud rectangular/square

**Dial Scale Marking Color:**

Black single indicator single range

**Accuracy In Percent:**

-1.0/+1.0 at full scale single indicator single range

**Meter Depth Behind Mounting Flange:**

0.960 inches

**Meter Body Diameter Behind Mounting Flange:**

2.690 inches

**Background Color:**

White single indicator single range

**Mounting Facility Type And Quantity:**

4 threaded stud

**Mounting Stud Diameter:**

0.138 inches

**Circuit Attachment Method And Quantity:**

2 threaded stud

**Style Designator:**

Rectangular flush mtg

**Fsc Application Data:**

Test set, elect. Equip.

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A310a0