

View Online at <https://aerobasegroup.com/nsn/6625-00-538-7074>

Indicator Type:

Spiral drum

Instrument Calibration Range:

8.20 to 12.4 gigahertz

Calibration Accuracy Rating:

Dial accuracy, porm 0.050 pct; accuracy, porm 0.080 pct

Special Features:

Waveguide adapter; accuracy specified over 20 degrees c and 0 to 100 pct relative humidity

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

Yes - demil/mli

Fiig:

T134-h