

View Online at <https://aerobasegroup.com/nsn/6625-00-571-3270>

**Circuit Current For Which Designed:**

Dc

**Dielectric Withstanding Voltage In Volts:**

3000.0

**Overall Length:**

2.100 inches

**Mounting Hole Diameter:**

Between 0.143 inches and 0.157 inches

**Overall Diameter:**

3.510 inches

**Environmental Protection:**

Cold resistant and heat resistant and shock resistant and submersible and vibration resistant and watertight

**Mounting Bolt Circle Diameter:**

3.160 inches

**Furnished Items And Quantity:**

Mtg hdw 3; term. Hdw 2

**Indicator Type:**

Dial w/pointer

**Installation Design:**

Panel

**Mounting Arrangement Style:**

Three hole/stud bolt circle

**Sensitivity Rating:**

2000.0 ohms per volt

**Dial Scale Marking Color:**

White single indicator single range

**Accuracy In Percent:**

-3.0 to 3.0 at full scale single indicator single range

**Unable To Decode:**

Unable to decode or unable to decode

**Meter Depth Behind Mounting Flange:**

1.660 inches

**Meter Body Diameter Behind Mounting Flange:**

2.820 inches

**Movement Type:**

Moving coil, permanent magnet

**Background Color:**

Black single indicator single range

**Pointer Color:**

White single indicator single range

**Power Source:**

External

**Mounting Facility Type And Quantity:**

3 screw and 3 unthreaded hole

**Scale Graduation Type:**

Nonlinear single indicator single range

**Measurement Range:**

+0.0 to 40.0 volts, dc single indicator single range

**Features Provided:**

Luminous scale marking and magnetic shield and zero adjuster

**Movement Suspension Method:**

Jewel pivot base

**Circuit Attachment Method And Quantity:**

2 screw and 2 solder lug

**Special Features:**

Ruggedized

**Special Markings:**

White zone on scale shall be 26.6 to 29.4 volts

**Material:**

Stainless steel

**Nuclear Hardness Critical Feature:**

Nonhardened

**Surface Treatment:**

Enamel semigloss

**Special Test Features:**

Conforms to mil-m-10304/11

**Style Designator:**

Round flush mtg

**Supplementary Features:**

Calibration position shall be horizontal

**Fsc Application Data:**

Test set, elect. Equip.

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A310a0