NSN 6625-00-635-4307

Ammeter - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6625-00-635-4307

| Circuit Current For Which Designed: |
|---|
| Dc |
| Dielectric Withstanding Voltage In Volts: |
| 2000.0 |
| Overall Length: |
| 2.076 inches |
| Mounting Hole Diameter: |
| 0.156 inches |
| Scale Division Quantity: |
| 30 single indicator single range |
| Overall Diameter: |
| 3.500 inches |
| Mounting Bolt Circle Diameter: |
| 3.160 inches |
| Indicator Type: |
| Dial w/pointer |
| Installation Design: |
| Panel |
| Mounting Arrangement Style: |
| Three hole/stud bolt circle |
| Sensitivity Rating: |
| 30.0 microamperes |
| Dial Scale Marking Color: |
| Black single indicator single range |
| Accuracy In Percent: |
| -2.0/+2.0 at full scale single indicator single range |
| Meter Depth Behind Mounting Flange: |
| 1.704 inches |
| Meter Body Diameter Behind Mounting Flange: |
| 2.672 inches |
| Movement Type: |
| Moving coil, permanent magnet |
| Background Color: |
| White single indicator single range |
| Pointer Color: |
| Black single indicator single range |
| Power Source: |
| External |
| Mounting Facility Type And Quantity: |
| 3 unthreaded hole |
| |
| Scale Graduation Type: |
| Linear single indicator single range |

NSN 6625-00-635-4307

Ammeter - Page 2 of 2

A310a0



| Measurement Range: |
|---|
| -30.0/+30.0 microamperes, $$ dc single indicator single range |
| Scale Selection Method: |
| Terminal |
| Features Provided: |
| Zero adjuster |
| Circuit Attachment Method And Quantity: |
| 2 solder stud |
| Special Markings: |
| Batt adj at right of center scale |
| Nuclear Hardness Critical Feature: |
| Nonhardened |
| Style Designator: |
| Round flush mtg |
| Fsc Application Data: |
| Test set, electronic equipment |
| Shelf Life: |
| N/a |
| Unit Of Measure: |
| |
| Demilitarization: |
| No |
| Fiig: |