## NSN 6625-00-752-7958

Oscilloscope - Page 1 of 3



View Online at https://aerobasegroup.com/nsn/6625-00-752-7958 **Overall Length:** 17.000 inches **Overall Height:** 10.000 inches **Overall Width: 6.750** inches **Current Type:** Ac Frequency Rating: Between 60.0 hertz and 800.0 hertz Joint Electronics Type Designation System Item Name: Oscilloscope Joint Electronics Type Designation System Item Type Number: Type no. An/usm-898 **Accessory Components And Quantity:** Oscilloscope 1, fan base 1, test adapter 1, cathode ray tube filter 1, viewing hood 1, test prod 1, cathode ray tube scale 1 **Voltage Rating:** 110.0 volts and 117.0 volts and 124.0 volts and 220.0 volts and 240.0 volts **Beam Quantity: Graduation Interval:** Graduated in v/in. On both axes **Mounting Method:** Portable **Accuracy Range In Percent:** -3.0/+3.0 Screen Size: 3.000 in. **Sweep Circuits:** Included **Generator Driven Feature:** Not included **Sweep Circuit Type:** Horizontal **Sweep Frequency Range:** +0.200/+50000.000 hertz **Trigger Sweep Polarity Switch:** Included

Provided

Not provided

**Expanded Sweep:** 

**Amplifier Deflection Sensitivity Provision:** 

## NSN 6625-00-752-7958

Oscilloscope - Page 2 of 3



Vertical Deflection Sensitivity Voltage Rating:
14.0 millivolts rms per inch
Horizontal Deflection Sensitivity Voltage Rating:
2.121 volts rms per inch
Input Intensity Modulation Feature:
Not provided
Plate Direct Connection Deflection Sensitivity Provision:
Provided
Direct Connection Vertical Deflection Sensitivity Voltage Rating:
21.21 volts any acceptable outside diameter
Direct Connection Horizontal Deflection Sensitivity Voltage Rating:
26.694 volts any acceptable outside diameter
Frequency Response Rating:
Provided
Z-axis Frequency Response Rating:
Between 0.02 kilohertz and 100.0 kilohertz
Amplifier Input Impedance Rating Provision:
Provided
X-axis Input Impedance Rating:
100000.0 ohms
Y-axis Input Impedance Rating:
1.0 megohms
Z-axis Input Impedance Rating:
1.0 megohms
Amplitude Calibrating Voltage Waveform:
Square wave
Calibrating Peak To Peak Amplitude:
Between 50.0 millivolts and 100000.0 millivolts
Calibrated Screen Scale:
Included
Trace Storage Tube:
Not included
Camera Recording Device Accommodation:
Provided
Phase:
Single
Special Features:
X-axis frequency response rating: dc to 500 kc; y-axis frequency response rating: dc to 4 mc
Shelf Life:
N/a
Unit Of Measure:
Demilitarization:

No

## NSN 6625-00-752-7958

Oscilloscope - Page 3 of 3



Fiig:

T320-g