

View Online at <https://aerobasegroup.com/nsn/6625-00-815-5948>

Ac Voltage Rating:

Between 105.0 volts and 125.0 volts

Frequency Rating:

Between 50.0 hertz and 60.0 hertz

Major Components:

Amplifier; indicator; generator; noise measuring set; oscillator

Accessory Component Quantity:

2 and 1 and 1 and 1 and 1

Accessory Controlling Agency:

Philco corp

Accessory Identifying Number:

348-5264 part no. And 348-5297 part no. And 348-5010 part no. And 348-5079 part no. And 348-5292 part no.

Internal Battery Accommodation:

Not included

Inclosure Feature:

Multiple item w/o carrying case

Test Type For Which Designed:

Measurement of audio frequency levels; measurement of noise; measurement of distortion; measurement of multichannel signal source; measurement of line monitor

Operating Test Capability:

2line amplifier input porm 10 dbm output porm 20 cbm; volume lever indicator range 30 cps to 15000 cps terminating input minus 60 to plus 23 vu bridging input minus 40 to plus 23 vu; tone generator range 1000 cps porm 1 pct input lever 1.4 vrms output lever porm 2dbm; noise measuring set range audio to55000 cps; audio oscillator output range 20 cps to 15000 cps level minus 70 dbm to plus 34.7 dbm; signal generator distribution system range audio input lever plus 34.7 dbm maximum output lever plus 25 dbm

Phase:

Single

Functional Classification:

Aa-6.3

Functional Description:

Audio test bay consisting of 13 major components in a steel case. Provides facilities for testing and evaluating the operational status of audio frequency channels and associated equipments

Reference Data And Literature:

To 33a1-7-65-1

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T228-a