NSN 6625-00-829-0882

Oscilloscope Calibrator - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6625-00-829-0882

Ac Voltage Rating:
Between 115.0 volts and 230.0 volts
Frequency Rating:
Between 50.0 hertz and 60.0 hertz
Calibration Type:
Dc voltage
Instrument Calibration Range:
12.0 millivolts to 1.2 volts, with a 2.0 volt level provided to check linearity of a sampling bridge
Calibration Accuracy Rating:
Porm 0.25 percent accuracy
Shelf Life:
N/a
Unit Of Measure:
Demilitarization:
No
Fiig:
T263-b