# NSN 6625-01-140-0838

Instrument Shunt - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6625-01-140-0838

#### **Mounting Provision Quantity:**

2

### **Overall Length:**

5.250 inches

# Center To Center Distance Between Mounting Facilities Parallel To Length:

4.750 inches

# Center To Center Distance Between Mounting Facilities Parallel To Width:

0.750 inches

# **Overall Height:**

1.250 inches

# **Overall Width:**

1.500 inches

**Unthreaded Mounting Hole Diameter:** 

### 0.205 inches

# **End Application:**

# Shuntbox

Usage Design:

#### Internal

Element Material:

#### Copper

**Mounting Method:** 

Unthreaded hole

#### **Terminal Location:**

Тор

# Voltage Drop In Millivolts:

50.0

# Current Rating And Type:

30.0 amperes dc

# Criticality Code Justification:

Zzzy

# **Special Features:**

Item must comply with requirements of defense supply center columbus production standard no. Lo3201

# **Special Markings:**

50mv, 30 amp

# Terminal Type And Quantity:

4 screw

# Fsc Application Data:

Test set, electronic equipment

# Reference Number Differentiating Characteristics:

Areference number differentiating material will be in accordance with naval inventory control point activity hx quality control. Manufacturing

and testing specifications available at the dla icp

# Shelf Life:

# NSN 6625-01-140-0838

Instrument Shunt - Page 2 of 2

Unit Of Measure:

---

Demilitarization:

No

Fiig:

A33500

