

View Online at <https://aerobasegroup.com/nsn/6625-01-156-7950>

**Frequency Rating:**

Between 0.0 hertz and 18.0 gigahertz

**End Application:**

6695-01-081-0706 calibration set

**Indicator Type:**

Meter or oscilloscope

**Mounting Method:**

Coaxial connector

**Instrument Calibration Range:**

2 to 18 ghz

**Calibration Accuracy Rating:**

Dc to 8 ghz porm 0.06; 8 to 12.4 ghz porm 0.07; 12.4 to 18 ghz porm 0.10

**Criticality Code Justification:**

Bdhw

**Supplementary Features:**

Male type connector

**Fsc Application Data:**

Electrical, electronic properties measuring and test instruments

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

T263-b