

View Online at <https://aerobasegroup.com/nsn/6625-01-156-7951>

Frequency Rating:

Between 0.0 hertz and 18.0 gigahertz

End Application:

6695-01-081-0706 calibration set

Indicator Type:

Meter or oscilloscope

Mounting Method:

Coaxial connector

Instrument Calibration Range:

2 to 18 ghz

Calibration Accuracy Rating:

Dc to 12.4 ghz porm 0.12; 12.4 to 18 ghz porm 0.22

Criticality Code Justification:

Bdhw

Supplementary Features:

Male type connector

Fsc Application Data:

Electrical, electronic properties measuring and test instruments

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T263-b