

View Online at <https://aerobasegroup.com/nsn/6625-01-278-3349>

Ac Voltage Rating:

Between 90.0 volts and 125.0 volts

Frequency Rating:

Between 50.0 hertz and 60.0 hertz

End Application:

Capacitor p/n 3307320-1, pbq oxygen regulator tester p/n 3300223-6001

Scale Calibration:

0.00001 pf to 99, 999 f

Scaleaccuracy Rating In Percent:

-0.020/+0.020

Output Frequency Range:

+50.0/+60.0 hertz

Adjustment Method:

Variable capacitor

Phase:

Single

Supplementary Features:

This rlc digibridge tester (impedance bridge) used to ensure capacitance of 2.7 porm .3 pct microfarads & maximum dissipation factor of .3 pct

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

Yes - demil/mli

Fiig:

T228-d