

View Online at <https://aerobasegroup.com/nsn/6625-01-473-8876>

Current Type:

Ac

Frequency Rating:

Between 48.0 hertz and 440.0 hertz

End Application:

Aircraft radar test station (arts) an/gsm-345

Voltage Rating:

115.0 volts and 230.0 volts

Special Features:

Boption 94 includes a calibration certificate; option 95 includes a calibration report; item may also include options 05 and 11; option 05 is video waveform trigger system, option 11, probe power

Supplementary Features:

Must include firmware version v1.9/1.4. Version no v1.9/1.4 can be verified by performing the following steps on the oscilloscope: 1 power on scope; 2 depress front panel button "menu off/extended functions"; 3 depress softkey "cal/diag"; 4 depress softkey "ext diag"; 5 verify version no v1.9/1.4 on upper portion of screen

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

T320-g