

View Online at <https://aerobasegroup.com/nsn/6625-01-506-3626>

Ac Voltage Rating:

110.0 volts

Frequency Rating:

60.0 hertz

Calibration Type:

Variable output

Phase:

Single

Special Features:

Must be equal to the fluke 9500b/1100 o-scope calibrator w/ one 9530 150ps and 500ps pulse rise time and 3.2 ghz capable active head. Ac characteristic of operation shall include leveled sine to a of 1.1 ghz and rise time capability to less than 150ps. Shall be capable of upgrade to at least 6.4 ghz

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

Yes - demil/mli

Fiig:

T263-b