

View Online at <https://aerobasegroup.com/nsn/6625-01-506-3626>

**Ac Voltage Rating:**

110.0 volts

**Frequency Rating:**

60.0 hertz

**Calibration Type:**

Variable output

**Phase:**

Single

**Special Features:**

Must be equal to the fluke 9500b/1100 o-scope calibrator w/ one 9530 150ps and 500ps pulse rise time and 3.2 ghz capable active head. Ac characteristic of operation shall include leveled sine to a of 1.1 ghz and rise time capability to less than 150ps. Shall be capable of upgrade to at least 6.4 ghz

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

Yes - demil/mli

**Fiig:**

T263-b