NSN 6630-00-033-8778

Chloride Water Sampling Test Kit - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6630-00-033-8778

Component Quantity:

105

Nonsupply Items And Quantities:

4 oz bottle silver nitrate solution w/eyedropper 1; chloride 2 indicator powder pillow 100; measuring tube 1; mixing bottle 1; clippers 1; 6.25 x 5 1.5 in. Case 1

Special Features:

Chloride range 0-250 or 0-1000 mg/l; 1 drop equals 12.5 mg/l or 50 mgl as sodium chloride; procedure is based on mohr argentometric method; 100 tests

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fiig:

A23800