NSN 6630-00-047-7269

Titrator - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6630-00-047-7269

General Description:

Fiig: A23900

Measures residual chlorine or other halogens to .001 ppm, electrical requirements 115v ac, 50/60 hz, single phase-printed circuit board-agitator motor-transformer-potentiometer-switch-readout bandmeter-all enclosed in aluminum housing, accessories consist of-pipets-squeeze bottle-flow control valve-sample container-drip cup-mounting stand w/base-phenylarsene oxide 4 oz-buffers ph-4, ph-7-pottassium iodide-electrolyte tablets-dropping bottle, o/a dim incl stand-24 in. Hx8.625 in. Wx7.500 in. Deep

· · · · · · · · · · · · · · · · · · ·
Shelf Life:
N/a
Unit Of Measure:
Demilitarization:
No