

View Online at <https://aerobasegroup.com/nsn/6630-00-445-5482>

**Instrument Type For Which Designed:**

No instrument designation

**Industry Designated Name:**

Electrode, fluoide

**Special Features:**

Rmeasures fluoride down to 0.02 ppm, addition of tisab provides proper ph adjustment, approved by astm for theanalysis of water, drinking water, waste water, and seawater, method approved by epa, sm, and aoac for feeds

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A506n0