## NSN 6630-00-774-1179

Oil Viscosimeter Test Bath - Page 1 of 2



View Online at https://aerobasegroup.com/nsn/6630-00-774-1179

Accessory Components And Quantity:					
Thermo-regulator-1, ubbelohde viscometers-12					
Material And Location:					
Glass reservoir					
Heating Element Current Type:					
Ac					
Heating Element Phase:					
Single					
Heating Element Frequency In Hertz:					
60.0					
Heating Element Current Rating In Amps:					
8.0					
Heating Element Voltage In Volts:					
120.0					
Heating Element Wattage In Watts:					
1000.0					
Heating Method:					
Electrical					
Heating Element Type:					
Electric					
Element Quantity:					
2					
Pipette Quantity For Which Designed:					
6					
Operating Tempurature Indicating Liquid:					
Water and glycerin					
Constant Tempurature Control:					
Included					
Accuracy Rating:					
-0.04 degrees fahrenheit and 0.04 degrees fahrenheit					
-0.04 degrees fahrenheit and 0.04 degrees fahrenheit <b>Liquid Circulator:</b>					
Liquid Circulator:					
Liquid Circulator: Included					
Liquid Circulator: Included Reservoir Inside Diameter:					
Liquid Circulator: Included Reservoir Inside Diameter: 12.000 inches					
Liquid Circulator: Included  Reservoir Inside Diameter: 12.000 inches  Reservoir Inside Depth:					
Liquid Circulator: Included Reservoir Inside Diameter: 12.000 inches Reservoir Inside Depth: 12.000 inches					
Liquid Circulator: Included Reservoir Inside Diameter: 12.000 inches Reservoir Inside Depth: 12.000 inches Accommodated Pipette Design:					
Liquid Circulator: Included  Reservoir Inside Diameter: 12.000 inches  Reservoir Inside Depth: 12.000 inches  Accommodated Pipette Design: Ubbelohde					

N/a

## NSN 6630-00-774-1179

Oil Viscosimeter Test Bath - Page 2 of 2



		• •			
ш	ın	10	<i>(</i> )+	Measure:	

--

Demilitarization:

No

Fiig:

T398-b