

View Online at <https://aerobasegroup.com/nsn/6630-00-857-2279>

**General Description:**

Below standard white to above water white, color comparison range w/std furnished; astm method d-156; sample holding tube 16.5 min 17.5 maximum mm int diameter 20 in. Lg 0.125 in. Sm grad; color standard tube 16.5 min 17.5mm int diameter 19 in. Lg, color disks swivel mtd; ext lt source ac 110v 60w; w/separate color comparison chart; w/carry case wd; accessories 1 ea color table comparison chart, color standard half disk, 3 color standard whole disks

**Departure From Cited Document:**

As mod by test method standard 791 method 101.6 para 2 saybolt chromometer apparatus

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fiig:**

A23900