## NSN 6630-01-035-1975

Nitrogen Oxide Analyz Calibrator - Page 1 of 1



View Online at https://aerobasegroup.com/nsn/6630-01-035-1975

	_	
General	110000	'Intion:
General	DESCI	IDUIDII.

Pushbutton selection of 4 concentrations in 0.2-10 ppm range, porm 1 pct. Reproducibility, drift porm 1 pct/mo	n, flow rate 2-3 l/min; 5.25 in	۱.
H 17 in W 20 25 in D: 115 vac 50/60 ha 220 vac 50 hz		

H, 17 in. W, 20.25 in. D;	115 vac, 50/60 ha, 220 vac	, 50 hz
Shelf Life:		

N/a

Unit Of Measure:

--

Demilitarization:

No