

View Online at <https://aerobasegroup.com/nsn/6630-01-055-2700>

**End Application:**

Nha part number 65560400-9 coolanol flush station assembly, f15i aircraft support

**Test Type For Which Designed:**

Test levels of turbidity found in high quality filtered water

**Operating Test Capability:**

0 to 100 nephelometric turbidity units (ntu)

**Functional Description:**

Provides the sensitivity, stability and flexibility necessary to continuously track the low levels of turbidity found in high quality filtered water; equipped with the aquatre (r) interface; as many as 8 sepreate 1720d turbidimeters can be controlled by one operator with a single aquatrend interface

**Reference Data And Literature:**

Csto is33d7-80-17-1 fig 7-1/102

**Relationship To Similar Equipment:**

Replaces nsn 6630-01-056-0653 part number 1720 turbidimeter

**Supplementary Features:**

Manufacturers name: low range process turbidimeter with aquatred interface

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

T228-a