

View Online at <https://aerobasegroup.com/nsn/6630-01-055-2700>

End Application:

Nha part number 65560400-9 coolanol flush station assembly, f15i aircraft support

Test Type For Which Designed:

Test levels of turbidity found in high quality filtered water

Operating Test Capability:

0 to 100 nephelometric turbidity units (ntu)

Functional Description:

Provides the sensitivity, stability and flexibility necessary to continuously track the low levels of turbidity found in high quality filtered water; equipped with the aquatre (r) interface; as many as 8 sepreate 1720d turbidimeters can be controlled by one operator with a single aquatrend interface

Reference Data And Literature:

Csto is33d7-80-17-1 fig 7-1/102

Relationship To Similar Equipment:

Replaces nsn 6630-01-056-0653 part number 1720 turbidimeter

Supplementary Features:

Manufacturers name: low range process turbidimeter with aquatred interface

Shelf Life:

N/a

Unit Of Measure:

--

Demilitarization:

No

Fig:

T228-a