

View Online at <https://aerobasegroup.com/nsn/6630-01-070-5774>

**End Application:**

6630-00-222-3539

**Special Features:**

The instrument has a carrying handle on top, a retractable bale stand, the front panel contains a 3.0 in. X 4.5 in. Band indication meter, tube positioning control "on-off" switch with indicator light, and two calibration control knobs

**Functional Description:**

Instrument for rating the amount of heater tube deposits resulting from the testing of gas turbine fuel using ASTM method D-3241

**Special Test Features:**

The instrument uses a light reflective method i.e.: the greater the deposit, the higher the reading

**Supplementary Features:**

Rate, tube deposit, 115 vac, 60 hz, 6 x 8 x 8 inch, 3 foot power cord

**Shelf Life:**

N/a

**Unit Of Measure:**

--

**Demilitarization:**

No

**Fig:**

A23800